

# Dosing table Liquid



Litre solution	Concentration of the applied solution										
	0,25%	0,3%	0,5%	0,75%	1%	1,5%	2%	2,5%	3%	4%	5%
1 Litre	2,5 ml	3 ml	5 ml	7,5 ml	10 ml	15 ml	20 ml	25 ml	30 ml	40 ml	50 ml
2 Litre	5 ml	6 ml	10 ml	15 ml	20 ml	30 ml	40 ml	50 ml	60 ml	80 ml	100 ml
2,5 Litre	6,25 ml	7,5 ml	12,5 ml	18,75 ml	25 ml	37,5 ml	50 ml	62,5 ml	75 ml	100 ml	125 ml
3 Litre	7,5 ml	9 ml	15 ml	22,5 ml	30 ml	45 ml	60 ml	75 ml	90 ml	120 ml	150 ml
4 Litre	10 ml	12 ml	20 ml	30 ml	40 ml	60 ml	80 ml	100 ml	120 ml	160 ml	200 ml
5 Litre	12,5 ml	15 ml	25 ml	37,5 ml	50 ml	75 ml	100 ml	125 ml	150 ml	200 ml	250 ml
8 Litre	20 ml	24 ml	40 ml	60 ml	80 ml	120 ml	160 ml	200 ml	240 ml	320 ml	400 ml
10 Litre	25 ml	30 ml	50 ml	75 ml	100 ml	150 ml	200 ml	250 ml	300 ml	400 ml	500 ml
20 Litre	50 ml	60 ml	100 ml	150 ml	200 ml	300 ml	400 ml	500 ml	600 ml	800 ml	1000 ml
30 Litre	75 ml	90 ml	150 ml	225 ml	300 ml	450 ml	600 ml	750 ml	900 ml	1200 ml	1500 ml

## Application guidelines::

10 litres: 1 % applied solution = 100 ml disinfection agent + 9.9 litres of water

Prepare the water. Add the appropriate amount of the disinfection agent concentrate and stir briefly. The solution is ready to apply immediately. According to the RKI's recommendation, disinfection agent solutions should be changed each working day. The solution should be stored in clean and covered containers (instrument baths or storage caps).

